

ISO-KALIBRIERUNG LDL2/LDL4

Application	
Design	for types, LDL2, LDL4
Mechanical data	
Weight	[g] 10



ISO calibration certificate for conductivity sensors (initial calibration)

ISO-KALIBRIERUNG LDL2/LDL4

Remarks	
Remarks	the ISO calibration certificate is not recognized according to the EA MLA (Multilateral Agreement) and is therefore considered as not traced.
	The calibration certificate must be ordered in conjunction with a sensor.
	The corresponding sensor is delivered along with the calibration certificate.
	The calibration certificate is in two languages (German/English).
	Recalibration or subsequent orders are only possible if the device is returned in a clean condition, without any adaptations attached (condition on delivery) and accompanied by a completed Declaration of Decontamination.
	In the internet version of the data sheet a sample certificate is stored in the tab "Downloads".
Pack quantity	1 pcs.